

BASELINE / TRIBRACH LOG

CREW	DATE	PAGES
715	8/23 24 2011	1-3
716	12/8 2011	4 5 16
711	4/18 2012	7 8
713	5/9	9 10
714	6/21	11 12
712	10/16	13 14
715	11/21	15-17
711	10/14 2013	18, 19
715	12/3	20, 21
714	3/5 2014	22, 23
715	4/14	24, 25
712	10/13	26, 27
711	10/28 2015	28-30
716	10/28	31-32
711	5/7 2020	33-35

TRIBRACH CHECK LOG

DATE: 8-23-11

CREW: 715

TRIBRACH (S/N: / COT #) 30801

OPTICAL PLUMMET:

PASS ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:

PASS ADJUSTED _____

TRIBRACH (S/N: / COT #) 25299

OPTICAL PLUMMET:

PASS ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:

PASS ADJUSTED _____

TRIBRACH (S/N: / COT #) 711300

OPTICAL PLUMMET:

PASS _____ ADJUSTED NOT APPLICABLE _____

CIRCULAR VIAL:

PASS ADJUSTED _____

TRIBRACH (S/N: / COT #) _____

OPTICAL PLUMMET:

PASS _____ ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:

PASS _____ ADJUSTED _____

2011 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	GTS-605	START TIME:	6:15 AM	END TIME:		DATE:	AUG 24, 2011
S/N:	SS 0412	PARTY CHIEF:	FRANK VALENZUELA	CHAINMAN:	MATT KALE	OBSERVER:	SCOTT PRECHTEL
COT#	031379	INSTRUMENT:	STEVE SHORIE	WEATHER CONDITIONS:	CLEAR, CALM	WET BULB (°C) / RELATIVE HUMIDITY (%)	
PRISM OFFSET	-30mm	DATE:	8/23/2011	TOPOCON	TRIPLE	START	END
WEATHER METER	KESTREL 3500	TOPOCON	#1654072	DATE:	FEB 28, 2011	bulb=	19.9
S/N:	#1654072	TOPOCON	#655330	RH=	48.8	AVG	

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:		NOTES:				
			DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)	
1	0	150	150.003	150.004	1665	540'	1590	1290	
2	0	420	420.008	420.008	Temp (°C)	27.8	28.3	29.8	
3	0	1390	1390.025	1390.025	P (hPa)	920.4	920.5	920.6	
4	150	1390	1240.019	1240.021	HI @ (150)	1662	1290	1600	1586
5	150	420	270.006	270.006	Temp (°C)	30.0	30.2	30.8	30.1
6	150	0	150.001	150.001	P (hPa)	920.8	920.8	920.8	921.0
7	420	0	420.005	420.004	HI @ (420)	1586	1577	1240	1516
8	420	150	270.004	270.003	Temp (°C)	31.4	31.3	31.3	32.0
9	420	1390	970.016	970.016	P (hPa)	921.0	921.2	921.3	921.8
10	1390	420	970.015	970.014	HI @ (1390)	1516	1516	1473	1516
11	1390	150	1240.020	1240.019	Temp (°C)	32.0	32.5	32.9	32.0
12	1390	0	1390.024	1390.022	P (hPa)	921.8	921.7	921.9	921.8

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

8/24/2011

ACQUISITION DATE: 8/24/11

PROCESS DATE:

FIELD CREW MEMBERS:

8/4/2010
715
Valenzuela
Schorie
Kale
Prechtel

TOPCON

Model#	GTS-605
S/N:	SS-0412
COT#	031379

DATA PROCESSOR:

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} *(m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0035	0.0020	0.0003
2	0	420	420.0090	420.0080	0.0010	-0.0005
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0060	-0.0024	-0.0040
6	150	0	150.0055	150.0010	0.0045	0.0028
7	420	0	420.0090	420.0045	0.0045	0.0030
8	420	150	270.0036	270.0035	0.0001	-0.0015
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000007 parts
 or S= -0.7 ppm
 C= 0.0018 meters

$\sigma_0^2 = 0.0000089$

$\sigma_s = 0.0000110$

$\sigma_c = 0.0033$

t (Student) Distribution

t _s =	0.064
t _c =	0.547

STATUS= **PASS**

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 1.82mm + -0.710ppm
 or 0.006ft + -0.710ppm

$D_{PUB} = D_{MEAS} + CF$

5

2011 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	GT5 233W	START TIME:	8:00 AM	END TIME:	10:15 AM	DATE:	Dec. 8, 2011
S/N:	261012	PARTY CHIEF:	RUSK	CHAINMAN:			
COT#	190263	INSTRUMENT:	10KE				
PRISM OFFSET	-30mm	WEATHER CONDITIONS:	COOL - CLEAR				
TRIBRACH CALIBRATION	DATE: KESTREL 3500	WET BULB (°C) / RELATIVE HUMIDITY (%)					
WEATHER METER	#1654072	bulb=	0.7	END	6.4		
S/N:	TOPCON TRIPLE	RH=	56.0	AVG	34.0		
PRISM MODEL#	TOPCON						
TRIBRACH CALIBRATION	DATE: FEB 28, 2011						
S/N:	TOPCON #655330						

NOTES:

OBS #	STATION:	FROM:	TO:	OBSERVED DISTANCE:	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
412	0	150	150	150.007	1.584	1418	1388	1321
211	0	420	420	420.010	Temp (°C)	11.2	12.5	12.9
210	0	1390	1390	1390.028	P (hPa)	925.3	925.4	925.3

OBS #	STATION:	FROM:	TO:	OBSERVED DISTANCE:	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
4A	150	1390	1390	1240.020	1.592	1321	1462	1468
58	150	420	420	270.007	Temp (°C)	9.9	8.9	7.7
61	150	0	0	150.006	P (hPa)	925.5	925.4	925.5

OBS #	STATION:	FROM:	TO:	OBSERVED DISTANCE:	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
46	420	0	0	420.009	1.613	11468	1509	1448
85	420	150	150	270.006	Temp (°C)	6.0	5.6	5.0
94	420	1390	1390	970.017	P (hPa)	925.4	925.2	925.4

OBS #	STATION:	FROM:	TO:	OBSERVED DISTANCE:	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
70	1390	420	420	970.015	1.524	1441	1507	1450
74	1390	150	150	1240.017	Temp (°C)	3.9	4.1	4.6
72	1390	0	0	1390.024	P (hPa)	925.9	925.9	925.9

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

2011
DEC, 2011

**BASELINE CALIBRATION TEST
ANGULAR MEASUREMENT CHECK**

5

OCCUPY: 150m

BACK SITE: 0m

FORE SITE: 420m

FORWARD

REVERSED

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 1:

0	0	0
179	59	50

180	00	00
359	59	54

DEGREE MINUTE SECOND

SET AVERAGE:

179	59	51.5
-----	----	------

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 2:

0	0	0
179	59	48

179	59	59
359	59	53

DEGREE MINUTE SECOND

SET AVERAGE:

179	59	51
-----	----	----

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 3:

0	0	0
179	59	50

180	00	06
359	59	57

DEGREE MINUTE SECOND

SET AVERAGE:

179	59	50.5
-----	----	------

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 4:

0	0	0
179	59	56

180	00	08
0	00	02

DEGREE MINUTE SECOND

SET AVERAGE:

179	59	55
-----	----	----

4 SET AVERAGE:

179	59	52
-----	----	----

OCCUPY: 150m

BACK SITE: 420m

FORE SITE :0m

FORWARD

REVERSED

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 1:

0	0	0
180	00	10

180	00	03
0	00	16

DEGREE MINUTE SECOND

SET AVERAGE:

180	00	11.5
-----	----	------

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 2:

0	0	0
180	00	09

180	00	01
0	00	15

DEGREE MINUTE SECOND

SET AVERAGE:

180	00	11.5
-----	----	------

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 3:

0	0	0
180	00	14

180	00	10
0	00	20

DEGREE MINUTE SECOND

SET AVERAGE:

180	00	12
-----	----	----

DEGREE MINUTE SECOND

DEGREE MINUTE SECOND

SET 4:

0	0	0
180	00	09

180	00	02
180	00	15

DEGREE MINUTE SECOND

SET AVERAGE:

180	00	11
-----	----	----

4 SET AVERAGE:

180	00	11.5
-----	----	------

CLOSING THE CIRCLE:

360	00	03.5
-----	----	------

02

23
12

MOD/SN: GTS-233W / 261012

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 12/8/11

PROCESS DATE: 12/8/2011

TOPCON

FIELD CREW MEMBERS:
Rugg
Luke
Prechtel

Model# GTS-233W
S/N: 261012
COT# 190263

DATA PROCESSOR: Prechtel

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0070	-0.0015	0.0002
2	0	420	420.0090	420.0100	-0.0010	0.0001
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0070	-0.0034	-0.0020
6	150	0	150.0055	150.0060	-0.0005	0.0012
7	420	0	420.0090	420.0090	0.0000	0.0011
8	420	150	270.0036	270.0055	-0.0019	-0.0005
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= 0.0000024 parts
or S= 2.4 ppm
C= -0.0020 meters

$\sigma_0^2 = 0.0000017$

$\sigma_s = 0.0000048$

$\sigma_c = 0.0014$

t (Student) Distribution

t_s= 0.491
t_c= 1.412

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = -2.05mm + 2.363ppm
or -0.007ft + 2.363ppm

$D_{PUB} = D_{MEAS} + CF$

12

2044-TOTAL STATION CALIBRATION LOG

1,458

TOPCON	MODEL#	GTS 233W	START TIME:	0800	END TIME:	10:42	DATE:	4/18/2012
	S/N:	261224	PARTY CHIEF:	STEVE ZEMMEL		CHAINMAN:	BOB DAN SULLIVAN	
	COT#	190454	INSTRUMENT:	SCOTT WEBER		OBSERVER:	SCOTT WEBER	
	PRISM OFFSET	-30mm	WEATHER CONDITIONS:	CLEAR - 0001 to warm		WET BULB (°C) / RELATIVE HUMIDITY (%)		
TRIBRACH CALIBRATION	DATE:	KESTREL 3500				START	END	AVG
WEATHER METER	S/N:	#1654072				bulb=	11.2	13.7
	TOPCON	TRIPLE				RH=	25.4	17.2
PRISM	MODEL#	TOPCON						
TRIBRACH CALIBRATION	DATE:	FEB 28, 2011						
	S/N:	TOPCON						
		#655330						

NOTES:

East.

Rept.

OBS #	STATION:	FROM:	TO:	DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150	150.005	150.005	1.647	1.376	1.306	1.448
2	0	420	420	420.009	420.009	Temp (°C)	22.7	23.8	24.2
3	0	1390	1390	1390.023	1390.023	P (hPa)	922.7	922.8	922.9

OBS #	STATION:	FROM:	TO:	DIRECT	REVERSED	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
4	150	1390	1390	1240.017	1240.019	1.669	1.448	1.472	1.458
5	150	420	420	270.004	270.004	Temp (°C)	23.9	24.4	25.1
6	150	0	0	150.002	150.001	P (hPa)	922.9	923.0	922.8

OBS #	STATION:	FROM:	TO:	DIRECT	REVERSED	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
7	420	0	0	420.007	420.007	1.549	1.458	1.502	1.518
8	420	150	150	270.005	270.002	Temp (°C)	26.6	26.6	27.8
9	420	1390	1390	970.011	970.012	P (hPa)	923.3	923.0	923.2

OBS #	STATION:	FROM:	TO:	DIRECT	REVERSED	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
10	1390	420	420	970.011	970.010	1.477	1.627	1.434	1.425
11	1390	150	150	1240.016	1240.015	Temp (°C)	27.8	28.3	28.0
12	1390	0	0	1390.022	1390.019	P (hPa)	923.2	923.6	923.5

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 4/18/12

PROCESS DATE: 4/18/2012

TOPCON

Model# GTS-233W
 S/N: 261224
 COT# 190454

FIELD CREW MEMBERS:
Ziemann
Weber
Sullivan

 DATA PROCESSOR:
Prechtel

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D_{PUB} (m)	D_{MEAS} (m)	Delta ($D_{PUB}-D_{MEAS}$)(m)	Residual V (m)
1	0	150	150.0055	150.0050	0.0005	-0.0013
2	0	420	420.0090	420.0090	0.0000	-0.0006
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0040	-0.0004	-0.0016
6	150	0	150.0055	150.0015	0.0040	0.0022
7	420	0	420.0090	420.0070	0.0020	0.0014
8	420	150	270.0036	270.0025	0.0011	-0.0001
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000043 parts
 or S= -4.3 ppm
 C= 0.0024 meters

$\sigma_0^2 = 0.0000029$

$\sigma_s = 0.0000063$

$\sigma_c = 0.0019$

t (Student) Distribution

$t_s = 0.676$
 $t_c = 1.261$

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	$t_{0.01}$
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 2.39mm + -4.262ppm
 or 0.008ft + -4.262ppm

$D_{PUB} = D_{MEAS} + CF$

6

2012
200912010 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	6TS-253W	START TIME:	6:55 AM	END TIME:		DATE:	5/9/2012
S/N:	263464	PARTY CHIEF:	JEFF WINGSTROM	CHAINMAN:			
COT#		INSTRUMENT:	BILL HENDERFADTH	OBSERVER:	SCOTT PROCKEL		
PRISM OFFSET	-30mm	WEATHER CONDITIONS:	2001-BREAZLY - cloudy	WET BULB (°C) / RELATIVE HUMIDITY (%)			
TRIBRACH CALIBRATION	KESTREL 3500	DATE:		START	END	AVG	
WEATHER METER	S/N: #1654072	TOPCON	TRIPLE	bulb=	13.8		
PRISM	MODEL#	DATE:	JUNE 02, 2009	RH=	38.0		
TRIBRACH CALIBRATION	S/N:	TOPCON	#655330				

NOTES:

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:		HI @ (0)			
			DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150.007	150.006	165.7	1.660	1.673	1.708
2	0	420	420.011	420.010	Temp (°C)	22.5	23.2	23.4
3	0	1390	1390.027	1390.025	P (hPa)	918.9	919.0	919.1
4	150	1390	1240.018	1240.019	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (1390)
5	150	420	270.006	270.005	Temp (°C)	23.4	24.2	24.6
6	150	0	150.003	150.004	P (hPa)	919.2	919.3	919.3
7	420	0	420.007	420.007	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
8	420	150	270.003	270.003	Temp (°C)	25.5	23.9	25.3
9	420	1390	970.016	970.015	P (hPa)	919.5	919.8	919.9
10	1390	420	970.014	970.016	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
11	1390	150	240.014	240.015	Temp (°C)	25.9	24.4	24.6
12	1390	0	1390.018	1390.020	P (hPa)	920.5	920.5	920.5

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

10

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE:

5/9/12

PROCESS DATE:

5/9/2012

FIELD CREW MEMBERS:

Wingstrom
Henerfauth

TOPCON

Model#

GTS-233W

S/N:

263464

COT#

DATA PROCESSOR:

Prechtel

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0065	-0.0010	-0.0011
2	0	420	420.0090	420.0105	-0.0015	-0.0014
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0	0.0019
5	150	420	270.0036	270.0055	-0.0019	-0.0019
6	150	0	150.0055	150.0035	0.0020	0
7	420	0	420.0090	420.0070	0.0020	0.0021
8	420	150	270.0036	270.0030	0.0006	0.0006
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000006 parts
 or S= -0.6 ppm
 C= 0.0002 meters

$\sigma_0^2 = 0.0000038$

$\sigma_s = 0.0000072$

$\sigma_c = 0.0022$

t (Student) Distribution

t_s= 0.089
 t_c= 0.098

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR =

0.21mm + -0.642ppm


or

0.001ft + -0.642ppm

D_{PUB}=D_{MEAS}+CF

2012 TOTAL STATION CALIBRATION LOG

(E)




TOPCON MODEL#	GTS-313	START TIME:	7:30 AM	END TIME:	10:30 AM	DATE:	June 21, 2012
S/N:	NY 1397	PARTY CHIEF:	LOFGREEN	CHAIRMAN:	LEHRLING	OBSERVER:	PRECHTEL
COT#		INSTRUMENT:	KALE	WEATHER CONDITIONS:	CLEAR	WET BULB (°C) / RELATIVE HUMIDITY (%)	
PRISM OFFSET	-30mm	DATE:		WEATHER CONDITIONS:	cool to hot	bulb=	14.9
TRIRBRACH CALIBRATION		DATE:		WEATHER CONDITIONS:		RH=	21.1
WEATHER METER		DATE:		WEATHER CONDITIONS:			18.4
S/N:	#1654072	DATE:		WEATHER CONDITIONS:			15.5
TRIRBRACH CALIBRATION		DATE:		WEATHER CONDITIONS:			
PRISM MODEL#	TRIPLE	DATE:		WEATHER CONDITIONS:			
TRIRBRACH CALIBRATION		DATE:		WEATHER CONDITIONS:			
S/N:	#655330	DATE:		WEATHER CONDITIONS:			

NOTES:

OBS #	STATION:	OBSERVED DISTANCE:		HI @ (0)			
		DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
12	0	150.009	150.009	1.664	5.20'	5.16'	5.25'
11	0	420.014	420.014	Temp (°C)	35.2	35.0	36.0
10	0	1390.032	1390.032	P (hPa)	920.0	920.0	920.0
9	150	1240.022	1240.019	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
8	150	270.004	270.003	Temp (°C)	33.6	33.7	32.4
7	150	150.004	150.004	P (hPa)	919.9	920.5	920.5
6	420	420.007	420.008	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
5	150	270.002	270.003	Temp (°C)	32.2	31.9	32.3
4	420	970.019	970.017	P (hPa)	920.7	920.7	920.8
3	1390	420.016	420.018	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
2	1390	1240.021	1240.023	Temp (°C)	30.4	31.3	30.5
1	1390	1390.027	1390.028	P (hPa)	921.5	921.3	921.1

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE



Call MJS
Hoof...

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 6/21/12

PROCESS DATE: 6/21/2012

TOPCON

Model# GTS-313
S/N: NY1397
COT#

FIELD CREW MEMBERS: Lofgreen
Kale
Lehrling

Prechtel

DATA PROCESSOR:

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0090	-0.0035	-0.0032
2	0	420	420.0090	420.0140	-0.0050	-0.0038
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0035	0.0001	0.0008
6	150	0	150.0055	150.0040	0.0015	0.0018
7	420	0	420.0090	420.0075	0.0015	0.0027
8	420	150	270.0036	270.0025	0.0011	0.0018
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000033 parts
or S= -3.3 ppm
C= 0.0002 meters

$\sigma_0^2 = 0.0000098$

$\sigma_s = 0.0000116$

$\sigma_c = 0.0035$

t (Student) Distribution

$t_s = 0.286$
 $t_c = 0.060$

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 0.21mm + -3.306ppm
or 0.001ft + -3.306ppm

$D_{PUB} = D_{MEAS} + CF$

2012 TOTAL STATION CALIBRATION LOG

13

TOPCON	MODEL#	GTS 233W	START TIME:	7:00 AM	END TIME:		DATE:	October 16, 2012
	S/N:	263198	PARTY CHIEF:	DAN ROBERTSON			CHAINMAN:	GRE ALSON
	COT#		INSTRUMENT:	"			OBSERVER:	SCOTT PEECHTEL
	PRISM OFFSET	-30mm	WEATHER CONDITIONS:	100% / CLEAR			WET BULB (°C) / RELATIVE HUMIDITY (%)	
TRIRACH CALIBRATION	DATE:	KESTREL 3500					START	END
WEATHER METER	S/N:	#1654072					bulb=	12.6
	MODEL#	TOPCON TRIPLE					RH=	32.3
TRIRACH CALIBRATION	DATE:	TOPCON						23.1
	S/N:	TOPCON #655330						

NOTES:

OBS #	STATION:	OBSERVED DISTANCE:		FROM:				TO:			
		DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150.004	150.003	1659	1544	1538	1613	1605	1478	1485	1485
2	0	420.008	420.008	Temp (°C)	15.0	15.5	16.5	22.0	20.6	22.0	22.0
3	0	1390.012	1390.013	P (hPa)	920.8	920.8	921.0	920.9	920.9	920.6	920.6
4	150	1240.010	1240.011	HI @ (150)	1.613	1.613	1.478	1.485	1.478	1.485	1.485
5	150	270.005	270.004	Temp (°C)	18.1	20.6	22.0	22.0	20.6	22.0	22.0
6	150	150.003	150.004	P (hPa)	921.0	920.9	920.9	920.6	920.9	920.6	920.6
7	420	420.006	420.006	HI @ (420)	1.664	1.664	1.469	1.557	1.469	1.557	1.557
8	420	270.003	270.003	Temp (°C)	21.9	22.2	22.2	23.5	22.2	22.2	23.5
9	420	1390.011	1390.012	P (hPa)	921.6	921.3	921.3	921.2	921.3	921.2	921.2
10	1390	420.008	420.009	HI @ (1390)	1.644	1.644	1.418	1.446	1.418	1.446	1.446
11	1390	1240.010	1240.013	Temp (°C)	23.8	24.4	24.4	24.8	24.4	24.4	24.8
12	1390	1390.014	1390.015	P (hPa)	921.6	921.7	921.7	921.5	921.7	921.5	921.5

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

2012
 10/19/12
 10:00

14

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 10/16/12

PROCESS DATE: 10/16/2012

TOPCON

Model# GTS-233W
S/N: 263198
COT#

FIELD CREW MEMBERS: Robertson
Rascon

DATA PROCESSOR: Prechtel

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0035	0.0020	0.0008
2	0	420	420.0090	420.0080	0.0010	-0.0004
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0045	-0.0009	-0.0022
6	150	0	150.0055	150.0035	0.0020	0.0008
7	420	0	420.0090	420.0060	0.0030	0.0016
8	420	150	270.0036	270.0030	0.0006	-0.0007
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= 0.0000006 parts
or S= 0.6 ppm
C= 0.0011 meters

$\sigma_0^2 = 0.0000023$

$\sigma_s = 0.0000056$

$\sigma_c = 0.0017$

t (Student) Distribution

t_s= 0.104
t_c= 0.661

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 1.12mm + 0.588ppm
or 0.004ft + 0.588ppm

D_{PUB}=D_{MEAS}+CF

TRIBRACH CHECK LOG

DATE: ~~11-21-12~~ 11-21-12

CREW: 715

TRIBRACH (S/N: / COT #) 711300

OPTICAL PLUMMET:
PASS ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:
PASS ADJUSTED _____

TRIBRACH (S/N: / COT #) 25299

OPTICAL PLUMMET:
PASS ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:
PASS ADJUSTED _____

TRIBRACH (S/N: / COT #) 30801

OPTICAL PLUMMET:
PASS ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:
PASS ADJUSTED _____

TRIBRACH (S/N: / COT #) _____

OPTICAL PLUMMET:
PASS _____ ADJUSTED _____ NOT APPLICABLE _____

CIRCULAR VIAL:
PASS _____ ADJUSTED _____

16

2012 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	ETS-605	START TIME:	9:00 AM	END TIME:	1:10 PM	DATE:	11-21-12
S/N:	550412	PARTY CHIEF:	Schorie	CHAINMAN:	Medina		
COT#	031379	INSTRUMENT:	Yates-Hodgson	OBSERVER:			
PRISM OFFSET	-30mm	WEATHER CONDITIONS:	Sunny, Cool	WET BULB (°C) / RELATIVE HUMIDITY (%)			
TRIBRACH CALIBRATION	DATE: 11-21-12			START	END	AVG	
WEATHER METER	S/N:			bulb=			
	KESTREL 3500			RH=			
	#1654072						
PRISM	MODEL#	TRIPLE					
TRIBRACH CALIBRATION	DATE: 11-21-12	FEB 28, 2011					
S/N:	TOPCON	#655330					

FROM: STATION: TO: STATION: OBSERVED DISTANCE: DIRECT REVERSED

OBS #	STATION	STATION	DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150.005	150.003	1546	1404	1444	1285
2	0	420	420.009	420.008	Temp (°C)	23.0	21.8	23.2
3	0	1390	1390.018	1390.016	P (hPa)	924.6	924.6	924.8

4	150	1390	1240.015	1240.016	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
5	150	420	270.007	270.005	1573	1285	1355	1365
6	150	0	150.007	150.007	Temp (°C)	23.5	25.3	25.6
					P (hPa)	924.8	924.6	924.4

7	420	0	420.007	420.008	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
8	420	150	270.004	270.006	1588	1365	1389	1228
9	420	1390	970.012	970.012	Temp (°C)	25.9	26.1	26.0
					P (hPa)	924.4	924.2	923.7

10	1390	420	970.016	970.015	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
11	1390	150	1240.021	1240.021	1468	1348	1331	1320
12	1390	0	1390.028	1390.023	Temp (°C)	27.2	26.3	27.8
					P (hPa)	924.0	923.6	923.6

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 11/21/12

PROCESS DATE: 11/21/2012

TOPCON

Model# GTS-605
S/N: SS-0412
COT# 031379

FIELD CREW MEMBERS: 715

Schorie
Yates-Hods
Medina

Prechtel

DATA PROCESSOR:

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0040	0.0015	0.0024
2	0	420	420.0090	420.0085	0.0005	0.0002
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0060	-0.0024	-0.0021
6	150	0	150.0055	150.0070	-0.0015	-0.0006
7	420	0	420.0090	420.0075	0.0015	0.0012
8	420	150	270.0036	270.0050	-0.0014	-0.0011
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= 0.0000043 parts
or S= 4.3 ppm
C= -0.0015 meters

$\sigma_0^2 = 0.0000032$

$\sigma_s = 0.0000066$

$\sigma_c = 0.0020$

t (Student) Distribution

t_s = 0.657
t_c = 0.762

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = -1.52mm + 4.344ppm
or -0.005ft + 4.344ppm

D_{PUB} = D_{MEAS} + CF

2013 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	GT5-233W	START TIME:	09:00	END TIME:	12:30	DATE:	10-14-2013	
S/N:	261224	PARTY CHIEF:	S Ziemann	CHAINMAN:	H Medina	OBSERVER:	D Luke	
COT#	190454	PRISM OFFSET	-30mm	WEATHER CONDITIONS:	Sunny, Cloud banks visible to South and East	WET BULB (°C) / RELATIVE HUMIDITY (%)		
TRIBRACH CALIBRATION	DATE: 10-14-2013	INSTRUMENT:	S Ziemann	WEATHER METER	KESTREL 3500	START	END	AVG
WEATHER METER	S/N: #1654072	TO: OBSERVED DISTANCE:	TO: STATION: DIRECT REVERSED	PRISM	MODEL# TRIPLE	14.56	15.0	14.78
TRIBRACH CALIBRATION	DATE: FEB 28, 2011	TO: STATION: DIRECT REVERSED	TO: STATION: DIRECT REVERSED	DATE:	TOPCON	41.0	29.3	35.15
S/N:	#65330	TO: STATION: DIRECT REVERSED	TO: STATION: DIRECT REVERSED	TOPCON	#65330			

NOTES:

OBS #	FROM: STATION:	TO: STATION:	DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150.0056	150.0052	1.607	1.550	1.522	1.290
2	0	420	420.0084	420.0076	Temp (°C)	22.0	22.9	23.1
3	0	1390	1390.0202	1390.0206	P (hPa)	924.3	924.2	924.0

OBS #	FROM: STATION:	TO: STATION:	DIRECT	REVERSED	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
4	150	1390	1240.0152	1240.0154	1.639	1.290	1.472	1.236
5	150	420	270.0038	270.0030	Temp (°C)	24.1	23.1	27.2
6	150	0	150.0056	150.0048	P (hPa)	924.3	924.1	923.9

OBS #	FROM: STATION:	TO: STATION:	DIRECT	REVERSED	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
7	420	0	420.0068	420.0070	1.639	1.236	1.401	1.336
8	420	150	270.0028	270.0022	Temp (°C)	24.1	23.9	25.7
9	420	1390	970.0132	970.0104	P (hPa)	923.8	923.7	923.5

OBS #	FROM: STATION:	TO: STATION:	DIRECT	REVERSED	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
10	1390	420	970.0126	970.0118	1.619	1.363	1.366	1.338
11	1390	150	1240.0182	1240.0178	Temp (°C)	25.6	27.0	26.2
12	1390	0	1390.0204	1390.0208	P (hPa)	923.8	923.7	923.5

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 10/14/13

PROCESS DATE: 10/14/2013

TOPCON

Model# GTS-233W
S/N: 261224
COT# 190454

FIELD CREW MEMBERS:
Ziemann
Luke
Medina

Prechtel

DATA PROCESSOR:

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0054	0.0001	0.0000
2	0	420	420.0090	420.0080	0.0010	-0.0005
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0034	0.0002	-0.0005
6	150	0	150.0055	150.0052	0.0003	0.0002
7	420	0	420.0090	420.0069	0.0021	0.0006
8	420	150	270.0036	270.0025	0.0011	0.0004
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= 0.0000050 parts
or S= 5.0 ppm
C= -0.0006 meters

$\sigma_0^2 = 0.0000003$

$\sigma_s = 0.0000019$

$\sigma_c = 0.0006$

t (Student) Distribution

t_s= 2.650
t_c= 1.068

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = -0.61mm + 5.041ppm

or -0.002ft + 5.041ppm

D_{PUB}=D_{MEAS}+CF

2013 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	GTS-605	START TIME:	9:30 AM	END TIME:	1:00 PM	DATE:	12/3/2013
S/N:	88-0412	PARTY CHIEF:	PETE - Y-H	CHAINMAN:	HERMAN MEDINA		
COT#	031379	INSTRUMENT:		OBSERVER:	SCOTT PRECHTEL		
PRISM OFFSET	-30mm	WEATHER CONDITIONS:	PARTLY CLOUDY & COOL		WET BULB (°C) / RELATIVE HUMIDITY (%)		
TRIBRACH CALIBRATION		DATE:			START	END	AVG
WEATHER METER		KESTREL 3500			bulb=	9.0	10.2
S/N:	#1654072				RH=	38.1	22.4
MODEL#	TRIPLE						
TRIBRACH CALIBRATION		DATE:					
S/N:	#655330						

NOTES:

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:		HI @ (0)			
			DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150.005	150.006	142.3 DM	134.7	128.0	129.4
2	0	420	420.009	420.0095	Temp (°C)	16.6	17.6	18.6
3	0	1390	1390.021	1390.0205	P (hPa)	918.8	919.1	919.2
4	150	1390	1240.011	1240.012	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
5	150	420	270.004	270.004	Temp (°C)	16.8	19.2	22.5
6	150	0	150.0055	150.0055	P (hPa)	918.9	918.9	918.5
7	420	0	420.010	420.0095	HI @ (420)	HR @ (0)	HR @ (150)	HR @ (1390)
8	420	150	270.004	270.004	Temp (°C)	19.8	20.2	20.0
9	420	1390	1970.01	970.011	P (hPa)	918.3	918.1	917.9
10	1390	420	970.014	970.0135	HI @ (1390)	HR @ (420)	HR @ (150)	HR @ (0)
11	1390	150	1240.017	1240.016	Temp (°C)	21.8	21.1	21.5
12	1390	0	1390.022	1390.022	P (hPa)	917.7	917.2	916.9

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

1871

PMS Sample #33

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 12/3/13

PROCESS DATE: 12/3/2013

TOPCON

FIELD CREW MEMBERS: 715

Model# GTS-605
S/N: SS-0412
COT# 031379

Yates-Hods
Medina

DATA PROCESSOR:

Prechtel

*** 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS**

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0055	0.0000	0.0001
2	0	420	420.0090	420.0093	-0.0003	0.0003
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0040	-0.0004	-0.0001
6	150	0	150.0055	150.0055	0.0000	0.0001
7	420	0	420.0090	420.0098	-0.0008	-0.0002
8	420	150	270.0036	270.0040	-0.0004	-0.0001
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000020 parts
or S= -2.0 ppm
C= 0.0002 meters

$\sigma_0^2 = 0.0000000$

$\sigma_s = 0.0000007$

$\sigma_c = 0.0002$

t (Student) Distribution

t_s = 2.723
t_c = 1.097

STATUS= **PASS**

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 0.24mm + -1.994ppm
or 0.001ft + -1.994ppm

D_{PUB} = D_{MEAS} + CF

2014 TOTAL STATION CALIBRATION LOG

NIKON
TOPCON MODEL# NIVD 3.14
 SN: A151805
 COT# 192844
 PRISM OFFSET -30mm
 TRIBRACH CALIBRATION DATE: 3/4/2014
 WEATHER METER KEISTREL 3500
 S/N: #1654072
 TRIBRACH CALIBRATION TOPCON TRIPLE
 S/N: #655330

START TIME: 7:15 AM	END TIME: 9:45 AM	DATE: MARCH 5, 2014
PARTY CHIEF: TATE LOFGREEN		CHAINMAN: NICK LEHRLING
INSTRUMENT: SCOTT WEBER		OBSERVER: SCOTT PRECHTEL
WEATHER CONDITIONS: CLEAR, COOL		
WET BULB (°C) / RELATIVE HUMIDITY (%)		
bulb=	START	END
RH=	11.4	12.3
	49.4	46.6
		48.0
		AVG

OBS #	STATION:	FROM:		TO:		OBSERVED DISTANCE:		TEMPERATURE:		PRESSURE:	
		STATION:	DIRECT	REVERSED	STATION:	DIRECT	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150.006	150.006	150	150.006	150.006	16.03	1.637	1.590	1.595
2	0	420	420.011	420.010	420	420.011	420.011	Temp (°C)	12 (12.0)	13.0 (12.2)	15.0
3	0	1390	1390.016	1390.016	1390	1390.016	1390.016	P (hPa)	923.50	923.60	924.0
4	150	1390	1240.012	1240.011	1390	1240.012	1240.012	HI @ (150)	1.575	1.540	1.477
5	150	420	270.006	270.005	420	270.006	270.006	Temp (°C)	15. (1.1)	17 (16.6)	16 (15.6)
6	150	0	150.007	150.008	0	150.007	150.007	P (hPa)	924 (60)	924 (58)	924 (70)
7	420	0	420.012	420.012	0	420.012	420.012	HI @ (420)	1.550	1.524	1.533
8	420	150	270.005	270.004	150	270.005	270.005	Temp (°C)	16 (16.5)	16 (15)	17 (16)
9	420	1390	970.010	970.011	1390	970.010	970.010	P (hPa)	924 (62.0)	924 (63)	924 (13)
10	1390	420	970.013	970.014	420	970.013	970.013	HI @ (1390)	1.555	1.549	1.558
11	1390	150	1240.012	1240.011	150	1240.012	1240.012	Temp (°C)	19 (1.2)	18 (1.4)	19 (1.2)
12	1390	0	1390.021	1390.023	0	1390.021	1390.021	P (hPa)	925 (1.0)	925 (4.7)	924 (7)

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
 ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

NIVD TS
 will not report
 Temp & Pressure
 in Precision

2.8
 2.0
 1.2

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 3/5/14

PROCESS DATE: 3/5/2014

NIKON
Model# NIVO 3 M
S/N: A151805
COT# 192844

FIELD CREW MEMBERS:
Lofgreen
Weber
Lehring

Prechtel

DATA PROCESSOR:

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0060	-0.0005	0.0006
2	0	420	420.0090	420.0105	-0.0015	0.0007
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0055	-0.0019	-0.0003
6	150	0	150.0055	150.0075	-0.0020	-0.0009
7	420	0	420.0090	420.0120	-0.0030	-0.0008
8	420	150	270.0036	270.0045	-0.0009	0.0007
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000038 parts
or S= -3.8 ppm
C= -0.0006 meters

$\sigma_0^2 = 0.0000007$

$\sigma_s = 0.0000031$

$\sigma_c = 0.0009$

t (Student) Distribution

t_s= 1.210
t_c= 0.609

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = -0.57mm + -3.784ppm

or -0.002ft + -3.784ppm

D_{PUB}=D_{MEAS}+CF

2014 TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	GTS 304	START TIME:	8:00 AM	END TIME:	11:30 AM	DATE:	APRIL 14, 2014
S/N:	FQ1105	PARTY CHIEF:	STEVE SCHORJE	CHAINMAN:			
COT#	-30mm	INSTRUMENT:	PETE YATES-HOPKINSON	OBSERVER:			
PRISM OFFSET		WEATHER CONDITIONS:	CLEAR/COOL	WET BULB (°C) / RELATIVE HUMIDITY (%)			
TRIBRACH CALIBRATION	KESTREL 3500			bulb=	11.6	END	21.5
WEATHER METER	#1654072			RH=	23.8	END	15.3
PRISM	TOPCON TRIPLE			NOTES:			
TRIBRACH CALIBRATION	DATE: FEB 28, 2011			27.4 23.4 109.8 137.0 248/96 79.96 1390 4170 4510			
S/N:	TOPCON #655330						

NOTES:

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:		HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
			DIRECT	REVERSED				
1	0	150	150.009	150.008	142.55	153.7	140.1	128.4
2	0	420	420.015	420.014	Temp (°C)	19.8	21.8	21.8
3	0	1390	1390.022	1390.020	P (hPa)	923.0	923.0	923.2
4	150	1390	1240.017	1240.011	HI @ (150)	123.4	149.6	149.1
5	150	420	270.007	270.008	Temp (°C)	22.8	22.9	24.1
6	150	0	150.011	150.009	P (hPa)	923.5	923.6	923.8
7	420	0	420.013	420.013	HI @ (420)	146.4	153.1	135.7
8	420	150	270.007	270.008	Temp (°C)	24.5	23.8	25.0
9	420	1390	970.016	970.013	P (hPa)	923.9	924.1	924.0
10	1390	420	970.014	970.019	HI @ (1390)	132.3	132.6	121.6
11	1390	150	1240.015	1240.021	Temp (°C)	24.3	24.1	24.0
12	1390	0	1390.027	1390.030	P (hPa)	924.5	924.5	924.5

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST

ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

24

Handwritten signature and date: 4/14/14

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 4/14/14

PROCESS DATE: 4/14/2014

TOPCON
Model# GTS-304
S/N: FQ1105
COT#

FIELD CREW MEMBERS: 715
Schorie
Yates-Hods

Prechtel

DATA PROCESSOR:

*** 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS**

OBS #	FROM STATION	TO STATION	D_{PUB} (m)	D_{MEAS} (m)	Delta ($D_{PUB}-D_{MEAS}$)(m)	Residual V (m)
1	0	150	150.0055	150.0085	-0.0030	0.0006
2	0	420	420.0090	420.0145	-0.0055	-0.0008
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0075	-0.0039	0.0002
6	150	0	150.0055	150.0100	-0.0045	-0.0009
7	420	0	420.0090	420.0130	-0.0040	0.0007
8	420	150	270.0036	270.0075	-0.0039	0.0002
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000038 parts
or S= -3.8 ppm
C= -0.0031 meters

$\sigma_0^2 = 0.0000006$

$\sigma_s = 0.0000028$

$\sigma_c = 0.0009$

t (Student) Distribution

$t_s = 1.331$
 $t_c = 3.593$

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	$t_{0.01}$
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = -3.07mm + -3.784ppm
or -0.010ft + -3.784ppm

$D_{PUB} = D_{MEAS} + CF$

2013 TOTAL STATION CALIBRATION LOG

TOPCON	MODEL#	GTS-233W	START TIME:	08:25	END TIME:	10:14	DATE:	OCT 13 2014
	S/N:	263198	PARTY CHIEF:	BUDDI ULJEKUSK			CHAINMAN:	
	COT#	-30mm	INSTRUMENT:	DANIEL ROBERTSON			OBSERVER:	
	PRISM CALIBRATION		WEATHER CONDITIONS:	PRETTY NICE MAN			WET BULB (°C) / RELATIVE HUMIDITY (%)	
WEATHER METER	DATE:	KESTREL 3500						
	S/N:	#1654072						
PRISM	MODEL#	TOPCON	TRIPLE				bulb=	13.2
	TRIBRACH CALIBRATION	DATE:	FEB 28, 2011				RH=	33.7
	S/N:	TOPCON	#655330					14.9
								30.0

NOTES:

OBS #	FROM STATION:	TO STATION:	OBSERVED DISTANCE:		HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
			DIRECT	REVERSED				
1	0	150	150.005	150.005	1.683	1.624	1.599	1.553
2	0	420	420.008	420.008	Temp (°C)	22.2	24.0	23.2
3	0	1390	1390.010	1390.008	P (hPa)	924.6	924.4	924.5
4	150	1390	1240.002	1240.003	HI @ (150)	1.576	1.563	1.515
5	150	420	270.002	270.001	Temp (°C)	23.1	24.00	25.1
6	150	0	150.004	150.003	P (hPa)	924.6	924.7	924.5
7	420	0	420.007	420.006	HI @ (420)	1.647	1.700	1.658
8	420	150	270.003	270.002	Temp (°C)	25.0	24.0	24.1
9	420	1390	970.000	969.999	P (hPa)	925.0	924.9	925.0
10	1390	420	970.002	970.001	HI @ (1390)	1.630	1.595	1.543
11	1390	150	1240.004	1240.003	Temp (°C)	25.0	25.3	26.3
12	1390	0	1390.008	1390.008	P (hPa)	925.6	925.7	925.6

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST

ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 10/13/14

PROCESS DATE: 10/13/2014

TOPCON
Model# GTS-233W
S/N: 263198
COT#

FIELD CREW MEMBERS:
Liljekvist
Robertson

Prechtel

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0050	0.0005	-0.0008
2	0	420	420.0090	420.0080	0.0010	-0.0008
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0015	0.0021	0.0006
6	150	0	150.0055	150.0035	0.0020	0.0007
7	420	0	420.0090	420.0065	0.0025	0.0007
8	420	150	270.0036	270.0025	0.0011	-0.0004
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6
S= 0.0000018 parts
or S= 1.8 ppm
C= 0.0010 meters

$\sigma_o^2 = 0.0000007$
 $\sigma_s = 0.0000031$
 $\sigma_c = 0.0009$

t (Student) Distribution
t_s = 0.591
t_c = 1.106

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 1.02mm + 1.817ppm
or 0.003ft + 1.817ppm D_{PUB} = D_{MEAS} + CF

TRIBRACH CHECK LOG

DATE: 10/28/15

CREW: 711

TRIBRACH (S/N: / COT #) S/N? 2152-03 (Chinese Manufacture)

OPTICAL PLUMMET:

PASS ADJUSTED ✓ NOT APPLICABLE

CIRCULAR VIAL:

PASS ✓ ADJUSTED

TRIBRACH (S/N: / COT #) Self-Serialized: #4570 ; "Sokkisha" Brand

OPTICAL PLUMMET:

PASS ✓ ADJUSTED NOT APPLICABLE

CIRCULAR VIAL:

PASS ✓ ADJUSTED

TRIBRACH (S/N: / COT #) No Numbers ; "GPS" marking ; New Chinese Manufacture

OPTICAL PLUMMET:

PASS ✓ ADJUSTED NOT APPLICABLE

CIRCULAR VIAL:

PASS ✓ ADJUSTED

TRIBRACH (S/N: / COT #)

OPTICAL PLUMMET:

PASS ADJUSTED NOT APPLICABLE

CIRCULAR VIAL:

PASS ADJUSTED

2015/2016 TOTAL STATION CALIBRATION LOG

711

TOPCON MODEL#	GTS-2330	START TIME:	8:10 AM	END TIME:	12:12 PM	DATE:	10/27/15
S/N:	261224	PARTY CHIEF:	Rugg + Ziemann	CHAINMAN:			
COT#	190450	INSTRUMENT:	Medina	OBSERVER:	Luke		
PRISM OFFSET	-30mm	WEATHER CONDITIONS:	Clear Light winds	WET BULB (°C) / RELATIVE HUMIDITY (%)			
TRIBRACH CALIBRATION	KESTREL 3500			bulb=	13.3	16.7	
WEATHER METER	#1654072			RH=	58.2	35.2	
S/N:							
PRISM MODEL#	TOPCON TRIPLE						
TRIBRACH CALIBRATION	DATE: OCT 27, 2015						
S/N:	TOPCON #655330						

NOTES:

OBS #	FROM: STATION:	TO: STATION:	REVERSED	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	150.0054	1.539	1,220	1,356	1,569
2	0	420	420.010	Temp (°C)	18.4	20.3	22.5
3	0	1390	1390.0174	P (hPa)	923.6	923.7	923.8
4	150	1390	1240.0076	HI @ (150)	1,569	1,355	1,474
5	150	420	270.0054	Temp (°C)	23.0	23.3	23.7
6	150	0	150.0054	P (hPa)	923.9	924.00	923.6
7	420	0	420.0074	HI @ (420)	1,474	1,219	1,567
8	420	150	270.0038	Temp (°C)	25.3	25.4	25.8
9	420	1390	970.0040	P (hPa)	923.7	923.6	923.5
10	1390	420	970.0040	HI @ (1390)	1,358	1,219	1,474
11	1390	150	1240.0084	Temp (°C)	27.2	27.3	27.2
12	1390	0	1390.0144	P (hPa)	924.0	924.0	923.6

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST
ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE:

10/27/15

PROCESS DATE:

10/28/2015

TOPCON

Model#

GTS-233W
261224
190454

S/N:

COT#

FIELD CREW MEMBERS:

Ziemann
Rugg
Medina
Luke

Prechtel

DATA PROCESSOR:

*** 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS**

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0053	0.0002	0.0003
2	0	420	420.0090	420.0098	-0.0008	0.0000
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0054	-0.0018	-0.0014
6	150	0	150.0055	150.0054	0.0001	0.0002
7	420	0	420.0090	420.0093	-0.0003	0.0005
8	420	150	270.0036	270.0036	0.0000	0.0004
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000024 parts

or S= -2.4 ppm

C= 0.0002 meters

$\sigma_0^2 = 0.0000006$

$\sigma_s = 0.0000029$

$\sigma_c = 0.0009$

t (Student) Distribution

t_s= 0.822
t_c= 0.270

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 0.24mm + -2.391ppm

or 0.001ft + -2.391ppm

D_{PUB}=D_{MEAS}+CF

2015/2016 TOTAL STATION CALIBRATION LOG

716

TOPCON	MODEL#	GTS - 233W	START TIME:	8:00AM	END TIME:	12:12 PM	DATE:	10/27/15
	S/N:	261012						
	COT#	190263	PARTY CHIEF:	Rugg + Zieman			CHAINMAN:	
	PRISM OFFSET	-30mm					OBSERVER:	Hermann
	TRIBRACH CALIBRATION		INSTRUMENT:	Leica				
WEATHER METER								
	S/N:	KESTREL 3500	WEATHER CONDITIONS:	Clear			WET BULB (°C) / RELATIVE HUMIDITY (%)	
	MODEL#	#1654072		Winds light to 3mph			bulb=	58.2
	TRIBRACH CALIBRATION						RH=	35.2
	S/N:							

NOTES:

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:		HI @ (150)	HR @ (150)	HR @ (420)	HR @ (1390)
			DIRECT	REVERSED				
1	0	150	150.008	150.008	1.538	1220	1356	1569
2	0	420	420.011	420.011	Temp (°C)	18.7	20.3	22.5
3	0	1390	1390.019	1390.018	P (hPa)	923.5	923.7	923.7
4	150	1390	1240.011	1240.010	HI @ (150)	1.569	1.355	1.474
5	150	420	270.006	270.006	Temp (°C)	23.2	20.5	23.3
6	150	0	150.005	150.005	P (hPa)	924.0	924.0	923.8
7	420	0	420.009	420.009	HI @ (420)	1.474	1.219	1.567
8	420	150	270.005	270.005	Temp (°C)	25.2	25.7	25.0
9	420	1390	970.003	970.003	P (hPa)	923.7	923.6	923.7
10	1390	420	970.005	970.006	HI @ (1390)	1.358	1.219	1.474
11	1390	150	1240.009	1240.009	Temp (°C)	27.6	27.5	27.3
12	1390	0	1390.012	1390.013	P (hPa)	924.1	923.8	923.6

RESET INSTRUMENT TO MEASURE IN METERS, °C & hPa AT THE BEGINNING OF THE CALIBRATION TEST

ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 10/27/15

PROCESS DATE: 10/28/2015

TOPCON
Model# GTS-233W
S/N: 261012
COT# 190263

FIELD CREW MEMBERS:
Rugg
Ziemann
Luke
Medina

Prechtel

DATA PROCESSOR:

*** 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS**

OBS #	FROM STATION	TO STATION	D_{PUB}^* (m)	D_{MEAS} (m)	Delta ($D_{PUB}-D_{MEAS}$)(m)	Residual V (m)
1	0	150	150.0055	150.0080	-0.0025	-0.0012
2	0	420	420.0090	420.0110	-0.0020	-0.0007
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0060	-0.0024	-0.0011
6	150	0	150.0055	150.0050	0.0005	0.0018
7	420	0	420.0090	420.0090	0.0000	0.0013
8	420	150	270.0036	270.0050	-0.0014	-0.0001
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= 0.0000002 parts
or S= 0.2 ppm
C= -0.0014 meters

σ_0^2 = 0.0000020
 σ_s = 0.0000053
 σ_c = 0.0016

t (Student) Distribution
t_s = 0.047
t_c = 0.866

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = -1.37mm + 0.246ppm
or -0.004ft + 0.246ppm $D_{PUB} = D_{MEAS} + CF$

TOTAL STATION CALIBRATION LOG

TOPCON	MODEL#	ES 102	START TIME:	11:45 AM	END TIME:	5/5/20
	S/N:	GS 0736	PARTY CHIEF:	S. Z.	CHAINMAN:	M. M.
	COT#		INSTRUMENT:	Topcon ES 102 V	OBSERVER:	D. M.
	PRISM OFFSET	-30mm	WEATHER CONDITIONS:	CLEAR, HOT	WET BULB (°C) / RELATIVE HUMIDITY (%)	
TRIBRACH CALIBRATION	DATE:	5/5/20	KESTREL 3500		START	END
WEATHER METER	S/N:	#1654072			18.6	18.3
	TOPCON	TRIPLE			16.1	14.0
PRISM	MODEL#				W bulb=	
TRIBRACH CALIBRATION	DATE:	5/5/20	TOPCON #655330		RH=	
	S/N:	May 29, 2008				

NOTES:

Head

HI @ (0)	1.565 m	HR @ (150)	1.492	HR @ (420)	1.525	HR @ (1390)	1.557
Temp (°C)			36.4		34.6		37.0
P (hPa)			921.0		921.3		921.5

HI @ (150)		HR @ (1390)		HR @ (420)		HR @ (0)	
Temp (°C)							
P (hPa)							

HI @ (420)		HR @ (0)		HR @ (150)		HR @ (1390)	
Temp (°C)							
P (hPa)							

HI @ (1390)		HR @ (420)		HR @ (150)		HR @ (0)	
Temp (°C)							
P (hPa)							

FROM:	TO:	OBSERVED
STATION:	STATION:	DISTANCE:
0	150	150.005
0	420	420.012
0	1390	1390.02

4	150	1390	meters
5	150	420	meters
6	150	0	meters

7	420	0	meters
8	420	150	meters
9	420	1390	meters

10	1390	420	meters
11	1390	150	meters
12	1390	0	meters

RESET INSTRUMENT TO MEASURE IN METERS AT THE BEGINNING OF THE CALIBRATION TEST
ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

TOTAL STATION CALIBRATION LOG

TOPCON MODEL#	ES 102	START TIME:	6:55 AM	END TIME:	9:40 AM	DATE:	5/6/20
S/N:	G-X0736	PARTY CHIEF:	S.Z.	INSTRUMENT:	Surveyor	CHAINMAN:	M, M,
COT#	-30mm	WEATHER CONDITIONS:	mild, clear				
PRISM OFFSET	DATE: 5/5/20	WET BULB (°C) / RELATIVE HUMIDITY (%)	START	END	AVG		
TRIBRACH CALIBRATION	KESTREL 3500		bulb=	14.3	21.7	18.0	
WEATHER METER	#1654072		RH=	32.0	27.2	29.6	
S/N:	TOPCON TRIPLE						
PRISM	DATE: 5/5/20						
TRIBRACH CALIBRATION	TOPCON #655330						
S/N:							

NOTES:

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:	HI @ (0)	HR @ (150)	HR @ (420)	HR @ (1390)
1	0	150	meters				
2	0	420	meters				
3	0	1390	meters				

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:	HI @ (150)	HR @ (1390)	HR @ (420)	HR @ (0)
4	150	1390	1240.012	1.554	1.552	1.575	
5	150	420	270.006	22.8	28.9	29.3	
6	150	0	150.004	922.7	921.9	921.8	

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:	HI @ (420)	HR @ (0)	HR (150)	HR @ (1390)
7	420	0	420.009	1.570	1.575	1.565	1.581
8	420	150	270.003	Temp (°C)	28.5	27.9	30.5
9	420	1390	970.010	P (hPa)	921.8	922.0	922.9

OBS #	FROM: STATION:	TO: STATION:	OBSERVED DISTANCE:	HI @ (1390)	HR @ (420)	HR (150)	HR @ (0)
10	1390	420	970.009	1.56	1.602	1.588	1.548
11	1390	150	1240.013	Temp (°C)	32.1	30.8	33.5
12	1390	0	1390.021	P (hPa)	922.2	922.2	921.9

RESET INSTRUMENT TO MEASURE IN METERS AT THE BEGINNING OF THE CALIBRATION TEST
ALL MEASUREMENTS CONVERTED FROM FEET TO METERS ARE UNACCEPTABLE

BASELINE CALIBRATION TEST LEAST SQUARES ANALYSIS

ACQUISITION DATE: 5/6/20

PROCESS DATE: 5/7/2020

TOPCON

FIELD CREW MEMBERS:

Model# ES-102
S/N: GX0736
COT#

Ziemann
Mullaney
Miranda

Prechtel

DATA PROCESSOR:

* 2004 NGS PUBLISHED DATA BETWEEN 0 & 420 MONUMENTS

OBS #	FROM STATION	TO STATION	D _{PUB} (m)	D _{MEAS} (m)	Delta (D _{PUB} -D _{MEAS})(m)	Residual V (m)
1	0	150	150.0055	150.0050	0.0005	-0.0002
2	0	420	420.0090	420.0120	-0.0030	-0.0013
3	0	1390	0.0000		0 0	
4	150	1390	0.0000		0 0	
5	150	420	270.0036	270.0060	-0.0024	-0.0020
6	150	0	150.0055	150.0040	0.0015	0.0008
7	420	0	420.0090	420.0090	0.0000	0.0017
8	420	150	270.0036	270.0030	0.0006	0.0010
9	420	1390	0.0000		0 0	
10	1390	420	0.0000		0 0	
11	1390	150	0.0000		0 0	
12	1390	0	0.0000		0 0	

N= 6

S= -0.0000090 parts
or S= -9.0 ppm
C= 0.0021 meters

s₀²= 0.0000026

s_s= 0.0000059

s_c= 0.0018

t (Student) Distribution

t_s= 1.523
t_c= 1.156

STATUS= PASS

"Degrees of Freedom" 99% Confidence level	
N-2	t _{0.01}
1	63.657
2	9.925
3	5.841
4	4.604
5	4.032
6	3.707
7	3.499
8	3.355
9	3.250
10	3.169
11	3.106
12	3.055

CORRECTION FACTOR = 2.07mm + -9.044ppm
or 0.007ft + -9.044ppm

D_{PUB}=D_{MEAS}+CF

LEAST SQUARES ANALYSIS USING ONLY THE DISTANCES BETWEEN THE 0, 150 AND 420 METER MONUMENTS